

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.		
10/750,579	12/29/2003	James Heggestuen	1001.1712101	8583		
28075 CROMPTON.	7590 10/10/2007 SEAGER & TUFTE, LLC		EXAM	EXAMINER		
1221 NICOLLI			DAWSON, GLENN K			
SUITE 800 MINNEAPOLIS, MN 55403-2420			ART UNIT	PAPER NUMBER		
			3731			
		,				
			MAIL DATE	DELIVERY MODE		
			10/10/2007	PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Advisory Action

Application No.	Applicant(s)	
10/750,579	HEGGESTUEN ET AL.	
Examiner	Art Unit	
Glenn K. Dawson	3731	

Before the Filing of an Appeal Brief -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --THE REPLY FILED 10 August 2007 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. 1. The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods: The period for reply expires _____months from the mailing date of the final rejection. b) 🛮 The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection. Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f). Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action, or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). NOTICE OF APPEAL 2. The Notice of Appeal was filed on __ ____. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a). **AMENDMENTS** 3. The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because (a) They raise new issues that would require further consideration and/or search (see NOTE below): (b) They raise the issue of new matter (see NOTE below): (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal: and/or (d) They present additional claims without canceling a corresponding number of finally rejected claims. NOTE: _____. (See 37 CFR 1.116 and 41.33(a)). 4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324). 5. Applicant's reply has overcome the following rejection(s): 6. Newly proposed or amended claim(s) ____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s). 7. Tor purposes of appeal, the proposed amendment(s): a) will not be entered, or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended. The status of the claim(s) is (or will be) as follows: Claim(s) allowed: Claim(s) objected to: Claim(s) rejected: Claim(s) withdrawn from consideration: AFFIDAVIT OR OTHER EVIDENCE 8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e). 9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1). 10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached. REQUEST FOR RECONSIDERATION/OTHER 11. 🛮 The request for reconsideration has been considered but does NOT place the application in condition for allowance because: See Continuation Sheet. 12. Note the attached Information Disclosure Statement(s). (PTO/SB/08) Paper No(s). 13. Other: ____. Glerin K Dawson

Primary Examiner Art Unit: 3731

Continuation of 11. does NOT place the application in condition for allowance because: The base reference teaches of the desirability to alter the curing of a stiffening material along the length of the catheter to alter the stiffness of the catheter along its length. The examiner contends that one skilled in the art would have found it obvious to use a material which could specifically alter the degree of curing of a polymeric material in order to alter the degree of curing and therefore alter the stiffness of respective portions of the catheter shaft, and that one attempting to do so would have found the teaching reference particularly relevant in that it teaches of adding to a polymeric material a photoinferter which only cures upon application of light, and that this allows for very specific degrees of curing of the polymeric material, therefore, the examiner contends that even though the teaching reference specifically deals with the adjustable curing of a polymeric dental article, that the examiner contends that this teaching could be applicable in virtually any polymeric material which needs to cure. The combination would allow the flexibility of the catheter to be intricately varied due to the degree of curing of the polymeric material. The exmainer contends that one skilled in the art would have therefore had every reason to expect to have success in providing a catheter of variable flexibility along it's length by adding a photoinferter to the polymeric material as it allows for intricate altering of the degree of curing of the material, which will affect the flexibility of the polymeric material making up the catheter.